

1. IDN:
`*IDN?\n`
2. Remote Ctrl Enable:
`SYSTem:REMOte\n`
3. Battery test mode:
`MODE BATTERY\n`
4. Set Current:
`BAT 2\n`
5. Ena cut-off voltage:
`BAT:VEN ON\n`
6. Cut-off voltage:
`BAT:VST 5\n`
7. Ena Battery Cap:
`BAT:CEN ON\n`
8. Discharge Cap:
`BAT:CST 2\n`
9. Ena Discharge Time:
`BAT:TEN ON\n`
10. DischargeTime Set:
`BAT:TIM 100\n`
11. Start test (Turn On):
`INP ON\n`
12. Stop test:
`INP OFF\n`
13. Pause test:
`INP PAUSE\n`
14. Test time query:
`BAT:TIME?\n`
15. Test State Query:
`BAT:STATe?\n`

16. Query all:

BAT:ALL?\n

In	Desc	SCPI Cmd
1	IDN:	*IDN?
2	Remote Ctrl Enable:	SYSTem:REMOte
3	Battery test mode:	MODE BATTERY
4	Set Current:	BAT 2
5	Ena cut-off voltage:	BAT:VEN ON
6	Cut-off voltage:	BAT:VST 5
7	Ena Battery Cap:	BAT:CEN ON
8	Discharge Cap:	BAT:CST 2
9	Ena Discharge Time:	BAT:TEN ON
10	DischargeTime Set:	BAT:TIM 100
11	Start test (Turn On):	INP ON
12	Stop test:	INP OFF
13	Pause test:	INP PAUSE
14	Test time query:	BAT:TIME?
15	Test State Query:	BAT:STATE?
16	Query all:	BAT:ALL?